

English

INSTRUCTIONS FOR USE: Southern Implants® PEEK Temporary Abutments



Southern Implants (Pty.) Ltd. 1 Albert Road, Irene, 0062, RSA
T: +27-12-667-1046 | E: info@southernimplants.com

EC REP

Southern Implants Europe AB: Holmgatan 30, S-791 71 Falun, Sweden
T: +46 23 13300 | E: ecrep@southernimplants.com

Subsidiaries

Australia

Southern Implants Australia
T: +61-(0)-8-9466-2627
E: info@southernimplants.com.au

Spain and Portugal

Southern Implants Iberica
T: +34 935 053 507
E: info@southernimplants.es

United Kingdom and Ireland

Southern Implants UK
T: +44-20-805-94490
E: info@southernimplants.co.uk

USA and Canada

Southern Implants North America Inc.
T: +1-561-472-0990
E: customercare@southernimplants.com

Description

The Southern Implants® PEEK Temporary Abutments are premanufactured and are available in a variety of connections, engaging and non-engaging, to fit the implant systems manufactured by Southern Implants. They are used as an aid in manufacturing a prosthesis for prosthetic rehabilitation. They can either be used for direct connection to an endosseous implant or they can be used for connecting the prosthesis to a compact conical abutment. The PEEK abutments are provided sterile; however, it will no longer be sterile after modification.

Intended use

Southern Implants PEEK Temporary Abutments are intended to be used in the Maxilla or Mandible for supporting a temporary prosthesis on endosseous implants in order to restore chewing function for the patient.

Indications for use

The Southern Implants PEEK Abutments are premanufactured prosthetic components directly connected to endosseous dental implants and are intended for provisional use up to 180 days as an aid in prosthetic rehabilitation.

Intended user

Dental Technicians, Maxillofacial Surgeons, General Dentists, Orthodontists, Periodontists, Prosthodontists, and other appropriately trained and/or experienced medical professionals.

Intended environment

The devices are intended to be used in a dental laboratory as part of the restoration design and manufacture as well as in a clinical environment such as an operating theatre or a dentist consultation room.

Intended patient population

The device is intended to be used in partially or fully edentulous patients eligible, or otherwise not contraindicated, for implant-retained prosthetic dental restorations.

Compatibility information

Southern Implants' implants should be restored with Southern Implants' components. In the Southern Implants' range there are 5 implant connections. The implant code and connection type can be identified by specific abbreviations in the product codes. Range identifiers are summarised in Table 1.

Table 1

Implant connection	Product code (non-radio opaque)	Product code (radio opaque)	Collar height	Compatible Platform	Engaging	Non engaging
External Hex	PKIP2H	PKR-EX-30	2mm	ø3mm External Hex	✓	
	PKIP2NH	PKR-NX-30	2mm	ø3mm External Hex		✓
	PKBN2H	PKR-EX-34	2mm	ø3.25mm External Hex	✓	
	PKBN2NH	PKR-NX-34	2mm	ø3.25mm External Hex		✓
	PKB2H	PKR-EX-40	2mm	ø4mm External Hex	✓	
	PKB2NH	PKR-NX-40	2mm	ø4mm External Hex		✓
	PKBA2H	PKR-EX-50	2mm	ø5mm External Hex	✓	
	PKBA2NH	PKR-NX-50	2mm	ø5mm External Hex		✓
	PKBBB2H	PKR-EX-60	2mm	ø6mm External Hex	✓	
	PKBBB2NH	PKR-NX-60	2mm	ø6mm External Hex		✓
	PKMAX9-2H	PKR-EX-70	2mm	ø9mm External Hex	✓	
	PKMAX9-2NH	PKR-NX-70	2mm	ø9mm External Hex		✓
	PKC-TOP1-EX-30	PKR-TOP1-EX-30	varying	ø3mm External Hex	✓	
	PKC-TOP1-EX-34	PKR-TOP1-EX-34	varying	ø3.25mm External Hex	✓	
	PKC-TOP1-EX-40	PKR-TOP1-EX-40	varying	ø4mm External Hex	✓	
	PKC-TOP2-EX-30	PKR-TOP2-EX-30	varying	ø3mm External Hex	✓	
PKC-TOP2-EX-34	PKR-TOP2-EX-34	varying	ø3.25mm External Hex	✓		

	PKC-TOP2-EX-40	PKR-TOP2-EX-40	varying	ø4mm External Hex	✓	
	PKC-TOP3-EX-34	PKR-TOP3-EX-34	varying	ø3.25mm External Hex	✓	
	PKC-TOP3-EX-40	PKR-TOP3-EX-40	varying	ø4mm External Hex	✓	
	PKC-TOP4-EX-34	PKR-TOP4-EX-34	varying	ø3.25mm External Hex	✓	
	PKC-TOP4-EX-40	PKR-TOP4-EX-40	varying	ø4mm External Hex	✓	
Tri-Nex	PKC-EL-35-2	PKR-EL-35	2mm	ø3.5mm Tri-nex	✓	
	PKC-NL-35-2	PKR-NL-35	2mm	ø3.5mm Tri-nex		✓
	PKC-EL-43-2	PKR-EL-43	2mm	ø4.3mm Tri-nex	✓	
	PKC-NL-43-2	PKR-NL-43	2mm	ø4.3mm Tri-nex		✓
	PKC-EL-50-2	PKR-EL-50	2mm	ø5mm Tri-nex	✓	
	PKC-NL-50-2	PKR-NL-50	2mm	ø5mm Tri-nex		✓
	PKC-EL-60-2	PKR-EL-60	2mm	ø6mm Tri-nex	✓	
	PKC-NL-60-2	PKR-NL-60	2mm	ø6mm Tri-nex		✓
Deep Conical (DC)	PKC-DC3-2	PKR-DC3	2mm	ø3mm Deep Conical	✓	
	PKC-NDC3-2	PKR-NDC3	2mm	ø3mm Deep Conical		✓
	PKC-DC4-2	PKR-DC4	2mm	ø4mm Deep Conical	✓	
	PKC-NDC4-2	PKR-NDC4	2mm	ø4mm Deep Conical		✓
	PKC-DC5-2	PKR-DC5	2mm	ø5mm Deep Conical	✓	
	PKC-NDC5-2	PKR-NDC45	2mm	ø5mm Deep Conical		✓
	PKC-TOP1-DC3	PKR-TOP1-DC3	varying	ø3mm Deep Conical	✓	
	PKC-TOP1-DC4	PKR-TOP1-DC4	varying	ø4mm Deep Conical	✓	
	PKC-TOP1-DC5	PKR-TOP1-DC5	varying	ø5mm Deep Conical	✓	
	PKC-TOP2-DC3	PKR-TOP2-DC3	varying	ø3mm Deep Conical	✓	
	PKC-TOP2-DC4	PKR-TOP2-DC4	varying	ø4mm Deep Conical	✓	
	PKC-TOP2-DC5	PKR-TOP2-DC5	varying	ø5mm Deep Conical	✓	
	PKC-TOP3-DC4	PKR-TOP3-DC4	varying	ø4mm Deep Conical	✓	
	PKC-TOP3-DC5	PKR-TOP3-DC5	varying	ø5mm Deep Conical	✓	
	PKC-TOP4-DC4	PKR-TOP4-DC4	varying	ø4mm Deep Conical	✓	
	PKC-TOP4-DC5	PKR-TOP4-DC5	varying	ø5mm Deep Conical	✓	
Internal Hex (M-series/ PROVATA)	PKC-3M-2	PKR-3M	2mm	ø3.3 Provata	✓	
	PKC-3NM-2	PKR-3NM	2mm	ø3.3 Provata		✓
	PKC-M-2	PKR-M	2mm	Internal Hex	✓	
	PKC-NM-2	PKR-NM	2mm	Internal Hex		✓
	PKC-Z-2	PKR-Z	2mm	Wide Internal Hex	✓	
	PKC-NZ-2	PKR-NZ	2mm	Wide Internal Hex		✓
Abutment level (Compact Conical)	PKC-MC	PKR-MC-48	1mm	ø4.8mm Compact Conical Abutments		✓
	PKC-MCW	PKR-MC-60	1mm	ø6mm Compact Conical Abutments		✓

Table 2 summarises the recommended screw torque values of the Southern Screws when used with the PEEK abutments.

Table 2

Range	Screw type	Torque
Abutment Level	1 Series (TSS1, TSU1, TSH1)	15 Ncm
External Hex (EX)	3 Series (TSSZ3, TSUZ3, TSHZ3)	< Ø4.0 mm implant interfaces: 15 Ncm ≥ Ø4.0 mm implant interfaces: 20 Ncm
	TS-P-16	15 Ncm
Tri-Nex (EL)	TS-L-18	15 Ncm
	TS-L-20	20 Ncm

Deep Conical (DC)	TS-DC3-14	15 Ncm
	TS-DC4-16	20 Ncm
	TS-DC5-20	20 Ncm
Internal Hex (M-Series & PROVATA®)	TS-Z-16	15 Ncm
	TS-Z-18	15 Ncm
Internal Octagon (IT)	TSIT2	20 Ncm

Clinical performance

The PEEK Abutments are intended to seal the implant connection and create a stable environment that fosters the integration of the implant into the surrounding bone and the healing of the soft tissue. Consequently, the clinical performance of the PEEK Abutments is primarily defined by their effectiveness in facilitating osseointegration and promoting optimal soft tissue healing. This performance can be quantitatively assessed through rates of successful osseointegration, the health of the peri-implant soft tissue, the incidence of soft tissue irritation, and the emergence profile established by the healing abutments.

Clinical benefits

Due to their role in the healing phase of dental implant therapy, the Southern Implants® PEEK Abutments provide clinical benefits associated with the enhancement of peri-implant soft tissue health and the improvement of aesthetic outcomes for the final restoration. Additionally, due to their intrinsic link with the overall success of the treatment, these abutments indirectly afford the benefits associated with the overall treatment, reflecting their contribution to the overall treatment system. These indirect benefits can be assessed through treatment success rates, patient satisfaction metrics and improvements in quality of life.

Storage, cleaning and sterilisation

The component is supplied sterile (sterilised by gamma irradiation) and intended for single-use prior to the expiration date (see packaging label). Sterility is assured unless the container or seal is damaged or opened. If packaging is damaged do not use the product and contact your Southern representative or return to Southern Implants®. The devices must be stored in a dry place at room temperature and not exposed to direct sunlight. Incorrect storage may influence device characteristics. Do not reuse components indicated for single-use only. Reusing these components may result in:

- damage to the surface or critical dimensions, which may result in performance and compatibility degradation.
- adds the risk of cross-patient infection and contamination if single-use items are reused.

Southern Implants® does not accept any responsibility for complications associated with reused single-use components.

Sterilisation

Southern Implants® recommends one of the following procedures to sterilise the restorations components prior to use:

1. prevacuum sterilisation method: steam sterilise the abutments at 132°C (270°F) at 180 - 220 kPa for 4 minutes. Dry for at least 20 minutes in the chamber. Only an approved wrap or pouch for steam sterilisation must be used.
2. for users in the USA: prevacuum sterilisation method: wrapped, steam sterilise at 135°C (275°F) at 180 - 220 kPa for 3 minutes. Dry for 20 minutes in the chamber. Use a wrap or pouch that is cleared for the indicated steam sterilisation cycle.

NOTE: users in the USA must ensure that the steriliser, wrap or pouch, and all steriliser accessories are cleared by the FDA, for the intended sterilisation cycle.

Contraindications

Do not use in patients:

- who are medically unfit for dental implant procedures.
- who are allergic or have hypersensitivity to polyetheretherketone (PEEK)
- where adequate numbers of implants could not be placed to achieve full functional support for a prosthesis

Warnings

THESE INSTRUCTIONS ARE NOT INTENDED AS A SUBSTITUTE FOR ADEQUATE TRAINING.

- For the safe and effective use of dental implants it is strongly suggested that specialised training be undertaken, including hands-on training to learn proper technique, biomechanical requirements and radiographic evaluations.
- Responsibility for proper patient selection, adequate training, experience in the placement of implants, and providing appropriate information for informed consent rests with the practitioner. Improper technique can result in implant failure, damage to nerves/vessels and/or loss of supporting bone.

Cautions

New and experienced Implant users should do training before using a new system or attempt to do a new treatment method.

- Take special care when treating patients who have local or systemic factors that could affect the healing of the bone and soft tissue. (I.e. poor oral hygiene, uncontrolled diabetes, are on steroid therapy, smokers, infection in the nearby bone and patients who had oro-facial radiotherapy.)
- Thorough screening of prospective implant candidates must be performed including:
 - A comprehensive medical and dental history.
 - Visual and radiological inspection to determine adequate bone dimensions, anatomical landmarks, occlusal conditions, periodontal status, and adequacy of bone.
 - Bruxism and unfavourable jaw relations must be taken into account.
- Proper pre-operative planning with a good team approach between well trained surgeons, restorative dentists and lab technicians is essential for successful implant treatment.
- Minimizing the trauma to the host tissue increases the potential for successful osseointegration.
- Electro-surgery should not be attempted around metal implants, as they are conductive
- The abutment level temporary abutments are only to be used for multi-unit restorations.

During surgery:

- Care must be taken that parts are not swallowed during any of the procedures, thus rubber-dam application is recommended when appropriate.
- Care must be taken to apply the correct tightening torque of abutments and abutment screws.

Post-surgery:

- Regular patient follow-up, and proper oral hygiene must be achieved are essential for favourable long-term results.

Procedures for use:

Chair side:

1. Shape the abutment to the correct occlusal height, with no additional angular correction. For single-unit restorations do not reduce to below a minimum height of 4mm from the top of the implant collar.
2. Connect the abutment to the implant with the correct prosthetic screw and driver (sold separately, refer to Table 2). Do not exceed the recommended torque.
3. Close the screw channel hole in a way that will ensure the prosthetic screw can be retrieved.
4. Make a temporary restoration by flowing a suitable temporary material such as acrylic resin into a pre-formed stent as described by Chu et al (Chu, S.J., Hochman, M.N., Tan-Chu, J.H., Mielezko, A.J. and Tarnow, D.P., 2014. A novel prosthetic device and method for guided tissue preservation of immediate postextraction socket implants. *Int J Periodontics Restorative Dent*, 34(Suppl 3), pp.s9-s17.).
5. Unscrew the temporary prosthesis.
6. Make final adjustments.
7. Clean and disinfect the restoration as applicable per the restorative material manufacturer's instructions.
8. Screw the temporary restoration back onto the implants and tighten the retaining screw to the correct torque values as indicated in Table 2.

Laboratory procedure:

The laboratory receives the impression from the clinician, either implant or abutment level.

1. The corresponding laboratory analogue is connected to the impression coping. Fabricate a working model with removable gingival mask or soft tissue material.
2. Follow the same steps 1-6 as for the clinical procedure to construct the temporary restoration.

Side effects

The clinical outcome of treatment is influenced by various factors. The following side effects and residual risks are associated with the device group and may necessitate further treatment, revision surgery or additional visits to the relevant medical professional's office. Furthermore, these side effects and residual risks can occur with varying possible severities and frequencies.

- Abutment fracture
- Allergic reaction(s) to the abutment material
- Bleeding
- Bruising
- Complications requiring revision surgery
- Discomfort
- Gingival inflammation
- Gingival recession
- Hematoma or bruising
- Hyperplastic tissue response
- Infection (acute and/or chronic)
- Localized inflammation
- Misfit or improper connection at the implant-abutment interface
- Overloading of the abutment/implant
- Pain or discomfort
- Pain or tenderness
- Peri-implantitis, peri-mucositis or otherwise poor peri-implant soft tissue health
- Periodontal inflammation
- Phonetic difficulties
- Soft tissue irritation
- Wound dehiscence or poor healing

Additionally, the normal side effects associated with anaesthesia should also be expected.

Precaution: maintaining sterility protocol

The abutments are packed in a peel pouch or blister base with a "peel-back" lid. Labelling information is located on the bottom half of the pouch, inside the packet or on the surface of the peel-back lid. Sterility is assured unless the pouch is damaged or opened.

Notice regarding serious incidents

Any serious incident that has occurred in relation with the device must be reported to the manufacturer of the device and the competent authority in the member state in which the user and/or patient is established.

The contact information for the manufacturer of this device to report a serious incident is as follows:
sicomplaints@southernimplants.com.

Materials

Abutment Material Type

White Medical Grade PEEK (VESTAKEEP® DC4420 R)

Abutment Screws Material Type

White Radio Opaque Medical Grade PEEK (VESTAKEEP® DC4430 R)

Grade 5 Titanium Alloy (ASTM F136 and Ti 6Al 4V)

CAT-8020-FDA-01 (01-2026)

CO-2113

Disposal

Disposal of the device and its packaging: follow local regulations and environmental requirements, taking different contamination levels into account. When disposing of spent items, take care of sharp drills and instruments. Sufficient PPE must be used at all times.

Magnetic Resonance (MR) Safety

This device has not been evaluated for safety and compatibility in the MR environment. It has not been tested for heating, migration or image artefact in the MR environment. The safety of this device in the MR environment is unknown. Scanning a patient who has this device may result in patient injury.

Summary of Safety and Clinical Performance (SSCP)

As required by the European Medical Device Regulation (MDR; EU2017/745), a Summary of Safety and Clinical Performance (SSCP) is available for perusal with regard to Southern Implants® product ranges.

The relevant SSCP can be accessed at <https://ec.europa.eu/tools/eudamed>.

NOTE: the above website will be available upon the launch of the European Database on Medical Devices (EUDAMED).

Disclaimer of liability

This product is part of the Southern Implants® product range and should only be used with the associated original products and according to the recommendations as in the individual product catalogues. The user of this product has to study the development of the Southern Implants® product range and take full responsibility for the correct indications and use of this product. Southern Implants® does not assume liability for damage due to incorrect use. Please note that some Southern Implants® products may not be cleared or released for sale in all markets.

Basic UDI

Product	Basic-UDI Number
Basic-UDI for PEEK Abutments	6009544038749A

Related literature and catalogues

CAT-2004 - Tri-Nex Implants Product Catalogue
 CAT-2005 - IT Implants Product Catalogue
 CAT-2020 - External Hex Implants Product Catalogue
 CAT-2042 - Deep Conical Implants Product Catalogue
 CAT-2043 - Internal Hex Implants Product Catalogue
 CAT-2060 - PROVATA® Implants Product Catalogue
 CAT-2069 - INVERTA® Implants Product Catalogue
 CAT-2070 - Zygomatic Implants Product Catalogue
 CAT-2092 - Soft Bone Implants Product Catalogue
 CAT-2093 - Single Platform Implants Product Catalogue
 CAT-2095 - External Hex INVERTA Implants Product Catalogue
 CAT-2096 - External Hex PTERYGOID Implants Product Catalogue

Symbols and warnings

											
Manufacturer: Southern Implants® 1 Albert Rd, P.O Box 605 IRENE, 0062, South Africa. Tel: +27 12 667 1046	CE mark 2797	Prescription device*	Sterilised using irradiation	Non-sterile	Use by date (mm-yy)	Do not reuse	Do not resterilise	Catalogue number	Batch code	Medical device	Authorised representative in the European Community
											
Authorised representative for Switzerland	Date of manufacture	Magnetic Resonance conditional	Magnetic Resonance safe	Single sterile barrier system with protective packaging inside	Single sterile barrier system	Consult instruction for use	Caution	Keep away from sunlight	Do not use if package is damaged		

* Prescription device: Rx only. Caution: Federal Law restricts this device to sale by or on the order of a licenced physician or dentist.

Canada licence exemption: Please note that not all products may have been licensed in accordance with Canadian law.

All rights reserved. Southern Implants®, the Southern Implants® logotype and all other trademarks used in this document are, if nothing else is stated or is evident from the context in a certain case, trademarks of Southern Implants®. Product images in this document are for illustration purposes only and do not necessarily represent the product accurately to scale. It is the responsibility of the clinician to inspect the symbols that appear on the packaging of the product in use.